






	<p>Do Not Use If Package is Damaged and consult instructions for use</p>	<p>Indicates that a medical device should not be used if the package has been damaged or opened and that the user should consult the instructions for additional information</p>	<p>Medical devices — Symbols to be used with information to be supplied by the manufacturer — Part 1: General requirements</p>	<p>ISO 15223-1 Reference #5.2.8</p>
	<p>Use by Date</p>	<p>Indicates the date after which the medical device is not to be used.</p>	<p>Medical devices — Symbols to be used with information to be supplied by the manufacturer — Part 1: General requirements</p>	<p>ISO 15223-1 Reference #5.1.4</p>
	<p>Date of Manufacture</p>	<p>Indicates the date when the medical device was manufactured.</p>	<p>Medical devices — Symbols to be used with information to be supplied by the manufacturer — Part 1: General requirements</p>	<p>ISO 15223-1 Reference #5.1.3</p>
	<p>Number of Products</p>	<p>Custom symbol denotes quantity per box.</p>	<p>NA</p>	<p>NA</p>
	<p>Keep dry</p>	<p>Indicates a medical device that needs to be protected from moisture.</p>	<p>Medical devices — Symbols to be used with information to be supplied by the manufacturer — Part 1: General requirements</p>	<p>ISO 15223-1 Reference #5.3.4</p>
	<p>Single sterile barrier system with protective packaging inside</p>	<p>Indicates a single sterile barrier with protective packaging inside</p>	<p>Medical devices — Symbols to be used with information to be supplied by the manufacturer — Part 1: General requirements</p>	<p>ISO 15223-1 Reference #5.2.13</p>
	<p>Do Not Resterilize</p>	<p>Indicates a medical device is not to be resterilized</p>	<p>Medical devices — Symbols to be used with information to be supplied by the manufacturer — Part 1: General requirements</p>	<p>ISO 15223-1 Reference #5.2.6</p>
	<p>Unique Device Identifier</p>	<p>Indicates a carrier that contains unique device identifier information</p>	<p>Medical devices — Symbols to be used with information to be supplied by the manufacturer — Part 1: General requirements</p>	<p>ISO 15223-1 Reference #5.7.10</p>